

6-MAM																																							
Sample	Target	Lab 1	Lab 2	Lab 3	Lab 4	Lab 5	Lab 6	Lab 7	Lab 8	Lab 9	Lab 10	Lab 11	Lab 12	Lab 13	Lab 14	Lab 15	Lab 16	Lab 17	Lab 18	Lab 19	Lab 20	Lab 21	Lab 22	Lab 23	Lab 24	Lab 25	Lab 26	Mean	Your Lab	%Diff	StDev	Mean -2SD	Mean +2SD	Mean -30%	Mean +30%	Within 2SD?			
PS1		0.0	0.0						0.0	0.0	0.0	0.0	0.0	0.0	0.0				0.0	0.0	0.0						0.0	0.0	0.0										
PS2		0.0	0.0						0.0	0.0	0.0	0.0	0.0	0.0	0.0				0.0	0.0	0.0						0.0	0.0	0.0										
PS3		0.0	0.0						0.0	0.0	0.0	0.0	0.0	0.0	0.0				0.0	0.0	0.0						0.0	0.0	0.0										
PS4		0.0	0.0						0.0	0.0	0.0	0.0	0.0	0.0	0.0				0.0	0.0	0.0						0.0	0.0	0.0										
PSS		0.0	0.0						0.0	0.0	0.0	0.0	0.0	0.0	0.0				0.0	0.0	0.0						0.0	0.0	0.0										
QCA	240	301.5	133.7						>200		288.7	362.0	465.5	285.7	321.0			302.5		242.2	273.4						266.5	294.8	301.5	2.3%	80.3	134.1	455.5	206.4	383.2	TRUE			
QCB	160	290.7	134.5						>200		252.0	368.0	398.5	221.7	329.0			294.4		199.1	233.8						248.4	270.0	290.7	7.7%	76.5	117.1	422.9	189.0	351.0	TRUE			
QCC	80	136.2	55.6						106.5		116.6	148.1	168.5	119.6	168.4			145.0		112.3	119.7						116.9	126.1	136.2	8.0%	30.7	64.8	187.4	88.3	164.0	TRUE			
QCD	40	70.3	28.3						52.3		57.4	56.6	79.8	47.2	81.7			74.9		59.6	61.1						61.1	60.8	70.3	15.5%	14.9	31.1	90.6	42.6	79.1	TRUE			
QCE	12	<25	9.4						15.8		16.0	21.9	27.4	14.6	23.8			19.6		18.2	18.9						19.1	25.0	25.0	0.0%	4.8	15.4	34.6	17.5	32.5	TRUE			

Notes:

R ²	Slope	Ave/Abs	Within 2SD?
0.999	1.048	6.69%	TRUE

